

LABORATOIRE DE
CONTROLE QUALITE

Responsable
Manager

J.FOSSE

Techniciens
Technicians

M. DUFOUR

L. FORTRIE

S. MONSAURET

T. LEURIDAN

C. CARON

H.DEPOORTER

S.BAILLEUL

A.MOREIRA

T. DELGHEIER

CERTIFICAT D'ANALYSE ANALYSIS CERTIFICATE

PRODUIT : *Bandelettes de contrôle
pour Anioxy-Twin*

PRODUCT *Anioxy Twin control strips*

LOT : 205022

BATCH NUMBER

Date d'analyse : 05/06/2020

Analysis date

Vérification de l'efficacité des bandelettes à l'aide de solutions d'acide peracétique ajustées à 850 et à 900 ppm.

Verification of strips' efficacy using 850 and 900 ppm peracetic acid solutions.

Virage non homogène (traces ou points blancs) à 850 ppm en 10 secondes effectué sur 10 bandelettes.

No colour, quick colour disappearance (small points or coloured edge), within 10 seconds performed with 10 strips.

Virage homogène au noir à 900 ppm en 10 secondes effectué sur 10 bandelettes.

Homogenous positive answer at 900 ppm (dark colouring) within 10 seconds performed with 10 strips.

PRODUIT CONFORME
PRODUCT CONFORMITY

OUI
YES

NON
NO

Date d'émission
Sending out date
05/06/2020

Visa Technicien Qualité
Quality Technician Visa



Visa Responsable Contrôle Qualité
Quality Control Manager Visa

