



Abbott Diabetes Care Limited  
 Range Road  
 Witney  
 Oxon OX29 0YL

Tel: 0870 400 2600  
 Fax: 0870 400 2800

**CERTIFICATE OF CONFORMITY AND APPROVAL FOR RELEASE**

**FAMILY** : G3C15  
**PRODUCT NAME** : FREESTYLE OPTIUM BLOOD GLUCOSE TEST STRIPS  
**PACK SIZE** : 100  
**RE-ORDER / CATALOGUE NUMBER** : 71505-75  
**LOT NUMBER** : 4500182654  
**MANUFACTURE DATE** : 2019/10/10  
**EXPIRY DATE** : 2021/03/31  
**NUMBER OF PACKS FOR RELEASE** : 11687

**CALIBRATION DATA**

**PRODUCT TEST DATA**

**INTERCEPT:** 0.133  $\mu\text{C}$  **SLOPE:** 0.00629  $\mu\text{C}/\text{mg/dL}$   
 0.11328  $\mu\text{C}/\text{mmol/L}$

**EXPECTED RESULTS FOR USE WITH CONTROLS**

**ASSIGNED RANGES:**

**TARGET:**

<b>HIGH CONTROL</b>	:	216	-	366	291	mg/dL
		12.0	-	20.3	16.1	mmol/L
<b>LOW CONTROL</b>	:	31	-	61	46	mg/dL
		1.7	-	3.4	2.5	mmol/L
<b>MID CONTROL</b>	:	68	-	124	96	mg/dL
		3.8	-	6.9	5.3	mmol/L

The above product has been manufactured in accordance with Abbott Diabetes Care Operating Procedures and conforms in all respects with the specific finished goods specifications as tested and inspected by Product Test, and conforms to order part numbers as defined by the Quality System.

**AUTHORISED BY :** John Whittaker

**DATE :** 24-Oct-2019 12:05

Registered Office in UK:  
 Abbott Diabetes Care Limited  
 Range Road, Witney,  
 Oxon OX29 0YL  
 Registered No. 2548046  
 Val Reg No. GB 537 1507 52

