



Abbott Diabetes Care Limited
 Range Road
 Witney
 Oxon OX29 0YL

Tel: 08 00 400 2000
 Fax: 0870 400 2000

CERTIFICATE OF CONFORMITY AND APPROVAL FOR RELEASE

FAMILY : G3CHP
 PRODUCT NAME : FREESTYLE OPTIUM NEO H BLOOD GLUCOSE TEST STRIPS
 PACK SIZE : 100
 RE-ORDER / CATALOGUE NUMBER : 71312-75
 LOT NUMBER : 4500183233
 MANUFACTURE DATE : 2020/02/26
 EXPIRY DATE : 2021/07/31
 NUMBER OF PACKS FOR RELEASE : 11615

PRODUCT TEST DATA

CALIBRATION DATA

INTERCEPT:	0.107	µC	SLOPE:	0.00645	µC/mg/dL
				0.11621	µC/mmol/L

EXPECTED RESULTS FOR USE WITH CONTROLS

ASSIGNED RANGES:				TARGET:	
HIGH CONTROL	:	240	-	389	315 mg/dL
		13.3	-	21.6	17.5 mmol/L
LOW CONTROL	:	34	-	65	50 mg/dL
		1.9	-	3.6	2.8 mmol/L
MID CONTROL	:	72	-	128	100 mg/dL
		4.0	-	7.1	5.6 mmol/L

The above product has been manufactured in accordance with Abbott Diabetes Care Operating Procedures and conforms in all respects with the specific finished goods specifications as tested and inspected by Product Test, and conforms to order part numbers as defined by the Quality System.

AUTHORISED BY : Mark Brown

DATE : 19-Mar-2020 10:41

Registered Office in UK
 Abbott Diabetes Care Limited
 Range Road, Witney,
 Oxon OX29 0YL
 Registered No. 2848546
 Vat Reg No. GB 537 1507 57

