



Abbott Diabetes Care Limited  
Range Road  
Witney  
Oxon OX29 0YL

Tel: 0870 400 2600  
Fax: 0870 400 2800

**CERTIFICATE OF CONFORMITY AND APPROVAL FOR RELEASE**

**FAMILY** : G3CHP  
**PRODUCT NAME** : FREESTYLE PRECISION PRO BLOOD GLUCOSE TEST STRIPS  
  
**PACK SIZE** : 100  
**RE-ORDER / CATALOGUE NUMBER** : 71195-70  
**LOT NUMBER** : 8A1AP0  
**MANUFACTURE DATE** : 2020/09/18  
**EXPIRY DATE** : 2022/02/28  
**NUMBER OF PACKS FOR RELEASE** : 11314

**PRODUCT TEST DATA**

**CALIBRATION DATA**

**INTERCEPT:** 0.122  $\mu\text{C}$  **SLOPE:** 0.00629  $\mu\text{C}/\text{mg}/\text{dL}$   
0.11328  $\mu\text{C}/\text{mmol}/\text{L}$

**FULL BARCODE NUMBER:** 8A1AP0uut8bx03

**EXPECTED RESULTS FOR USE WITH CONTROLS**

**ASSIGNED RANGES:**

**TARGET:**

<b>HIGH CONTROL</b>	:	231	-	381	306	<b>mg/dL</b>
		12.8	-	21.1	17.0	<b>mmol/L</b>
<b>LOW CONTROL</b>	:	32	-	62	47	<b>mg/dL</b>
		1.8	-	3.4	2.6	<b>mmol/L</b>
<b>MID CONTROL</b>	:	69	-	125	97	<b>mg/dL</b>
		3.8	-	6.9	5.4	<b>mmol/L</b>

The above product has been manufactured in accordance with Abbott Diabetes Care Operating Procedures and conforms in all respects with the specific finished goods specifications as tested and inspected by Product Test, and conforms to order part numbers as defined by the Quality System.

**AUTHORISED BY :** Mark Brown

**DATE :** 19-Oct-2020 10:35

Registered Office in UK:  
Abbott Diabetes Care Limited  
Range Road, Witney,  
Oxon OX29 0YL  
Registered No. 2548046  
Vat Reg No. GB 537 1507 52

