

LABORATOIRE DE  
CONTROLE QUALITE

Responsable  
Manager

J.FOSSE

Techniciens  
Technicians

M. DUFOUR

L. FORTRIE

S. MONSAURET

T. LEURIDAN

A.MOREIRA

C.CARON

H. DEPOORTER

S.BAILLEUL

T. DELGHEIER

**CERTIFICAT D'ANALYSE**  
ANALYSIS CERTIFICATE

**PRODUIT** : *Bandelettes de contrôle*

**LOT** : 202017

*pour Stéranios (2%-2%NG-2%ECS)* BATCH NUMBER

**PRODUCT** *Stéranios control strips*

Date d'analyse : 04/06/2020

Analysis date

Vérification de l'efficacité des bandelettes à l'aide de solutions de Glutaraldéhyde ajustées à 1,5%, 1,8% et à 2,0%.

Verification of strips' efficacy using 1,5% 1,8% and 2,0% Glutaraldehyde solutions.

Contrôle du virage à la couleur rose orange (1,5%) en 15 secondes effectué sur 10 bandelettes.

Verification of negative answer at 1,5% (Pink orange colouring within 15 seconds) performed with 10 strips.

Contrôle du virage au rose uniforme (1,8%) en 15 secondes effectué sur 10 bandelettes.

Verification of positive answer at 1,8% (Evenly pink colouring within 15 seconds) performed with 10 strips.

Contrôle du virage au rose uniforme (2,0%) en 15 secondes effectué sur 10 bandelettes.

Verification of positive answer at 2,0% (Evenly pink colouring within 15 seconds) performed with 10 strips.

**PRODUIT CONFORME**  
PRODUCT CONFORMITY

**OUI**   
YES

**NON**   
NO

Date d'émission  
Sending out date  
04/06/2020

Visa Technicien Qualité  
Quality Technician Visa



Visa Responsable Contrôle Qualité  
Quality Control Manager Visa

