



Abbott Diabetes Care Limited
 Range Road
 Witney
 Oxon OX29 0YL

Tel: 0870 400 2600
 Fax: 0870 400 2900

CERTIFICATE OF CONFORMITY AND APPROVAL FOR RELEASE

FAMILY : G3CHP
PRODUCT NAME : FREESTYLE OPTIUM NEO H BLOOD GLUCOSE TEST STRIPS

PACK SIZE : 100
RE-ORDER / CATALOGUE NUMBER : 71312-75
LOT NUMBER : 4500184625
MANUFACTURE DATE : 2021/02/18
EXPIRY DATE : 2022/07/31
NUMBER OF PACKS FOR RELEASE : 11357

PRODUCT TEST DATA

CALIBRATION DATA

INTERCEPT:	0.122	µC	SLOPE:	0.00591	µC/mg/dL
				0.10645	µC/mmol/L

EXPECTED RESULTS FOR USE WITH CONTROLS

ASSIGNED RANGES:				TARGET:		
HIGH CONTROL	:	229	-	378	304	mg/dL
		12.7	-	21.0	16.9	mmol/L
LOW CONTROL	:	27	-	58	43	mg/dL
		1.5	-	3.2	2.4	mmol/L
MID CONTROL	:	65	-	121	93	mg/dL
		3.6	-	6.7	5.2	mmol/L

The above product has been manufactured in accordance with Abbott Diabetes Care Operating Procedures and conforms in all respects with the specific finished goods specifications as tested and inspected by Product Test , and conforms to order part numbers as defined by the Quality System.

AUTHORISED BY : Steven Sherriff

DATE : 23-Mar-2021 11:47

Registered Office in UK:
 Abbott Diabetes Care Limited
 Range Road, Witney,
 Oxon OX29 0YL
 Registered No 2548046
 Vat Reg No GB 537 1507 52

