

LABORATOIRE DE
CONTROLE QUALITE

Responsable
Manager

J.FOSSE

Techniciens
Technicians

M. DUFOUR

L. FORTRIE

S. MONSAURET

T. LEURIDAN

A.MOREIRA

C.CARON

H. DEPOORTER

S.BAILLEUL

T. DELGHEIER

CERTIFICAT D'ANALYSE
ANALYSIS CERTIFICATE

PRODUIT : *Bandelettes de contrôle*

LOT : 202016

pour Stéranios (2%-2%NG-2%ECS) BATCH NUMBER

PRODUCT *Stéranios control strips*

Date d'analyse : 07/01/2020

Analysis date

Vérification de l'efficacité des bandelettes à l'aide de solutions de Glutaraldéhyde ajustées à 1,5%, 1,8% et à 2,0%.

Verification of strips' efficacy using 1,5% 1.8% and 2,0% Glutaraldehyde solutions.

Contrôle du virage à la couleur rose orange (1,5%) en 15 secondes effectué sur 10 bandelettes.

Verification of negative answer at 1,5% (Pink orange colouring within 15 seconds) performed with 10 strips.

Contrôle du virage au rose uniforme (1,8%) en 15 secondes effectué sur 10 bandelettes.

Verification of positive answer at 1,8% (Evenly pink colouring within 15 seconds) performed with 10 strips.

Contrôle du virage au rose uniforme (2,0%) en 15 secondes effectué sur 10 bandelettes.

Verification of positive answer at 2,0% (Evenly pink colouring within 15 seconds) performed with 10 strips.

PRODUIT CONFORME
PRODUCT CONFORMITY

OUI
YES

NON
NO

Date d'émission
Sending out date
19/10/2020

Visa Technicien Qualité
Quality Technician Visa



Visa Responsable Contrôle Qualité
Quality Control Manager Visa

