CERTIFICATE OF CONFORMITY AND APPROVAL FOR RELEASE

FAMILY	:	G3CHP
PRODUCT NAME	:	FREESTYLE PRECISION PRO BLOOD GLUCOSE TEST STRIPS
PACK SIZE	:	100
RE-ORDER / CATALOGUE NUMBER	:	71195-70
LOT NUMBER	:	8C3AP0
MANUFACTURE DATE	:	2021/11/26
EXPIRY DATE	:	2023/04/30
NUMBER OF PACKS FOR RELEASE	:	11573

PRODUCT TEST DATA

CALIBRATION DATA

INTERCEPT:	0.122	**µC**	**SLOPE:**	0.00629	**µC/mg/dL**
				0.11328	**µC/mmol/L**

FULL BARCODE NUMBER: 8C3AP0uuv79x03

EXPECTED RESULTS FOR USE WITH CONTROLS

ASSIGNED RANGES:					**TARGET:**	
HIGH CONTROL	:	231	-	381	306	**mg/dL**
		12.8	-	21.1	17.0	**mmol/L**
LOW CONTROL	:	30	-	60	45	**mg/dL**
		1.7	-	3.3	2.5	**mmol/L**
MID CONTROL	:	68	-	124	96	**mg/dL**
		3.8	-	6.9	5.4	**mmol/L**

The above product has been manufactured in accordance with Abbott Diabetes Care Operating Procedures and conforms in all respects with the specific finished goods specifications as tested and inspected by Product Test , and conforms to order part numbers as defined by the Quality System.

AUTHORISED BY : Mark Brown **DATE :** 13-Jan-2022 11:19

Abbott
A Promise for Life