CERTIFICATE OF CONFORMITY AND APPROVAL FOR RELEASE

FAMILY	:	G3CHP
PRODUCT NAME	:	FREESTYLE OPTIUM NEO H BLOOD GLUCOSE TEST STRIPS
PACK SIZE	:	100
RE-ORDER / CATALOGUE NUMBER	:	71312-75
LOT NUMBER	:	4500185731
MANUFACTURE DATE	:	2022/01/18
EXPIRY DATE	:	2023/06/30
NUMBER OF PACKS FOR RELEASE	:	11545

PRODUCT TEST DATA

CALIBRATION DATA

INTERCEPT:	0.122	**µC**	**SLOPE:**	0.00640	**µC/mg/dL**
				0.11523	**µC/mmol/L**

EXPECTED RESULTS FOR USE WITH CONTROLS

ASSIGNED RANGES:					**TARGET:**	
HIGH CONTROL	:	238	-	387	313	**mg/dL**
		13.2	-	21.5	17.4	**mmol/L**
LOW CONTROL	:	31	-	61	46	**mg/dL**
		1.7	-	3.4	2.6	**mmol/L**
MID CONTROL	:	70	-	126	98	**mg/dL**
		3.9	-	7.0	5.5	**mmol/L**

The above product has been manufactured in accordance with Abbott Diabetes Care Operating Procedures and conforms in all respects with the specific finished goods specifications as tested and inspected by Product Test , and conforms to order part numbers as defined by the Quality System.

AUTHORISED BY : John Whittaker **DATE :** 14-Feb-2022 12:36

Abbott
A Promise for Life