

CERTIFICATE OF CONFORMITY AND APPROVAL FOR RELEASE

FAMILY : G3CP
PRODUCT NAME : FREESTYLE OPTIUM BLOOD GLUCOSE TEST STRIPS

PACK SIZE : 50
RE-ORDER / CATALOGUE NUMBER : 71498-75
LOT NUMBER : 4500185840
MANUFACTURE DATE : 2022/02/18
EXPIRY DATE : 2023/07/31
NUMBER OF PACKS FOR RELEASE : 7934

PRODUCT TEST DATA

CALIBRATION DATA

INTERCEPT:	0.133	µC	SLOPE:	0.00629	µC/mg/dL
				0.11328	µC/mmol/L

EXPECTED RESULTS FOR USE WITH CONTROLS

ASSIGNED RANGES:	TARGET:
HIGH CONTROL : 216 - 366 12.0 - 20.3	291 mg/dL 16.1 mmol/L
LOW CONTROL : 31 - 61 1.7 - 3.4	46 mg/dL 2.5 mmol/L
MID CONTROL : 68 - 124 3.8 - 6.9	96 mg/dL 5.3 mmol/L

The above product has been manufactured in accordance with Abbott Diabetes Care Operating Procedures and conforms in all respects with the specific finished goods specifications as tested and inspected by Product Test , and conforms to order part numbers as defined by the Quality System.

AUTHORISED BY : John Whittaker

DATE : 11-Mar-2022 12:23