

CERTIFICATE OF CONFORMITY AND APPROVAL FOR RELEASE

FAMILY : G3CP  
PRODUCT NAME : FREESTYLE OPTIUM BLOOD GLUCOSE TEST STRIPS  
  
PACK SIZE : 100  
RE-ORDER / CATALOGUE NUMBER : 71505-75  
LOT NUMBER : 4500185159  
MANUFACTURE DATE : 2021/07/15  
EXPIRY DATE : 2022/12/31  
NUMBER OF PACKS FOR RELEASE : 5749

PRODUCT TEST DATA

CALIBRATION DATA

INTERCEPT: 0.133  $\mu\text{C}$  SLOPE: 0.00629  $\mu\text{C}/\text{mg}/\text{dL}$   
0.11328  $\mu\text{C}/\text{mmol}/\text{L}$

EXPECTED RESULTS FOR USE WITH CONTROLS

| ASSIGNED RANGES:         | TARGET:     |
|--------------------------|-------------|
| HIGH CONTROL : 216 - 366 | 291 mg/dL   |
| 12.0 - 20.3              | 16.1 mmol/L |
| LOW CONTROL : 31 - 61    | 46 mg/dL    |
| 1.7 - 3.4                | 2.5 mmol/L  |
| MID CONTROL : 68 - 124   | 96 mg/dL    |
| 3.8 - 6.9                | 5.3 mmol/L  |

The above product has been manufactured in accordance with Abbott Diabetes Care Operating Procedures and conforms in all respects with the specific finished goods specifications as tested and inspected by Product Test , and conforms to order part numbers as defined by the Quality System.

AUTHORISED BY : Michael King

DATE : 03-Aug-2021 9:31