

CERTIFICATE OF CONFORMITY AND APPROVAL FOR RELEASE

FAMILY : G3CHP  
PRODUCT NAME : FREESTYLE PRECISION PRO BLOOD GLUCOSE TEST STRIPS  
  
PACK SIZE : 100  
RE-ORDER / CATALOGUE NUMBER : 71195-70  
LOT NUMBER : 8C56P0  
MANUFACTURE DATE : 2022/01/10  
EXPIRY DATE : 2023/06/30  
NUMBER OF PACKS FOR RELEASE : 11685

PRODUCT TEST DATA

CALIBRATION DATA

|            |       |               |        |         |                                    |
|------------|-------|---------------|--------|---------|------------------------------------|
| INTERCEPT: | 0.122 | $\mu\text{C}$ | SLOPE: | 0.00618 | $\mu\text{C}/\text{mg}/\text{dL}$  |
|            |       |               |        | 0.11133 | $\mu\text{C}/\text{mmol}/\text{L}$ |

FULL BARCODE NUMBER: 8C56P0stv8bx01

EXPECTED RESULTS FOR USE WITH CONTROLS

| ASSIGNED RANGES: | TARGET: |      |   |      |      |        |
|------------------|---------|------|---|------|------|--------|
| HIGH CONTROL     | :       | 231  | - | 381  | 306  | mg/dL  |
|                  |         | 12.8 | - | 21.1 | 17.0 | mmol/L |
| LOW CONTROL      | :       | 32   | - | 62   | 47   | mg/dL  |
|                  |         | 1.8  | - | 3.4  | 2.6  | mmol/L |
| MID CONTROL      | :       | 69   | - | 125  | 97   | mg/dL  |
|                  |         | 3.8  | - | 6.9  | 5.4  | mmol/L |

The above product has been manufactured in accordance with Abbott Diabetes Care Operating Procedures and conforms in all respects with the specific finished goods specifications as tested and inspected by Product Test , and conforms to order part numbers as defined by the Quality System.

AUTHORISED BY : Steven Sherriff

DATE : 23-Feb-2022 13:36