

CERTIFICATE OF CONFORMITY AND APPROVAL FOR RELEASE

FAMILY

G3CHP

PRODUCT NAME

FREESTYLE PRECISION PRO BLOOD GLUCOSE TEST STRIPS

PACK SIZE

: 100

RE-ORDER / CATALOGUE NUMBER

71195-70

LOT NUMBER

8F14P0

MANUFACTURE DATE

2023/09/11

EXPIRY DATE

2025/02/28

NUMBER OF PACKS FOR RELEASE

11464

PRODUCT TEST DATA

CALIBRATION DATA

INTERCEPT:

0.122

μC

SLOPE:

0.00640

μC/mg/dL

0.11523

µC/mmol/L

FULL BARCODE NUMBER:

8F14P0wuu7cz08

EXPECTED RESULTS FOR USE WITH CONTROLS

ASSIGNED RANG	ES:				TARGET:	
HIGH CONTROL	1	235	+	385	310	mg/dL
		13.0		21.4	17.2	mmol/L
LOW CONTROL		33	-	63	48	mg/dL
		1.8	-	3.5	2.7	mmol/L
MID CONTROL	The second secon	71		127	99	mg/dL
*		3.9	-]	7.0	5.5	mmol/L

The above product has been manufactured in accordance with Abbott Diabetes Care Operating Procedures and conforms in all respects with the specific finished goods specifications as tested and inspected by Product Test , and conforms to order part numbers as defined by the Quality System.

AUTHORISED BY:

John Whittaker

DATE:

24-Oct-2023 13:53



